

IMP57

Rel 1.01 of 31/03/2005

Accessory for Loop/Line Impedance with high resolution

Page 1 of 2

1. ELECTRICAL SPECIFICATIONS

Accuracy is given as \pm (% of reading + no. of least significant digits) at 23 °C \pm 5 °C, with relative humidity less than 80% R.H.

Z _{LINE} Z _{LOOP} MEASUREMENT			
Range	Resolution	Accuracy	
$0.0 \div 199.9 \text{m}\Omega$	0.1 m Ω	$\pm (5.0\% \text{rdg+1m}\Omega)$	
$200 \div 1999 m\Omega$	1m Ω		

RLINE RLOOP, XLINE XLOOP MEASUREMENT		
Range	Resolution	Accuracy
$0.0 \div 199.9 \text{m}\Omega$	0.1 m Ω	\pm (10%rdg+2m Ω)
$200 \div 1999$ mΩ	1mΩ	

PROSPECTIVE SHORT CIRCUIT CURRENT			
Range	Resolution	Accuracy	
0 ÷ 1999A	1A		
2.0 ÷ 9.9kA	0.1kA	related to Z accuracy	
10 ÷ 1999kA	1kA		

VOLTAGE		
Range (50Hz ± 5%)	Resolution	Accuracy
190 ÷ 460V	1V	\pm (1.0%lettura+2cifre)

FREQUENCY		
Range (50Hz ± 5%)	Resolution	Accuracy
47.5 ÷ 52.5Hz	0.1Hz	±0.2Hz

HT ITALIA srl Via della Boaria 40 - 48018 Faenza (RA)- Italy Tel: +39-0546-621002 - Fax: +39-0546-621144 email: export@htitalia.it - web: http://www.htitalia.com



IMP57

Rel 1.01 of 31/03/2005

Accessory for Loop/Line Impedance with high resolution

Page 2 of 2

2. GENERAL SPECIFICATIONS

Measurement mode:

4 wire high resolution Line impedance measurement.

Function:

- Phase to Phase, Phase to Neutral, Phase to Protection Conductor Impedance Measurement.
- Phase to Phase, Phase to Neutral, Phase to Protection Conductor Prospective short circuit current Calculation.

Max Test Current:

202A

Conformity:

The instrument is according to: EN60909-0, VDE 0413, EN61557-3

Safety standards

The instrument complies with: EN 61010, EN61557 Insulation: Class 2, double insulation

Pollution level: 2 2000m Maximum height:

CAT III 240V (to Earth) Overvoltage category:

CAT III 415V (between P1, C1, P2, C2 inputs)

Power supply:

- 220 ÷ 415V nominal (from P1 and P2 inputs)
- Frequency: 50Hz±5%

Dimension:

340mm (W) x 300mm (L) mm x 150mm (H)

Weight:

approx. 4100g (without accessories).

Operating temperature:

0°C to 40 °C, <80 % R.H.

Storage temperature:

-10°C to 60°C, <80 % R.H.

Via della Boaria 40 - 48018 Faenza (RA)- Italy

HT ITALIA srl

Tel: +39-0546-621002 - Fax: +39-0546-621144 email: export@htitalia.it - web: http://www.htitalia.com